

Fast Analysis of Volatile Organic Compounds by Solid Phase Microextraction/Capillary GC

Solid phase microextraction is a fast, solventless alternative to conventional sample extraction techniques. Because no solvent is injected, and the analytes are rapidly desorbed onto the GC column, short, narrow bore capillary columns can be used. This greatly reduces analysis time and improves minimum detection limits, while maintaining resolution. SPME can be effective in analyses of volatile organic compounds, for sample screening or as an alternative extraction method in a formal analysis.

Key Words:

- volatile organic compounds • solid phase microextraction
- wastewater

In analyses of volatile organic compounds (VOCs) in water, a purge and trap system typically is used to concentrate the analytes. This methodology has important limitations. To efficiently remove the analytes from the trap, carrier gas flow rates greater than 5mL/minute are required. Long (60-105m), large bore (320-750 μ m) capillary columns are most compatible with these high flow rates, but such columns yield long analysis times (30-50 minutes). Without a jet separator, high flow rates are not compatible with a mass spectrometer, but a jet separator raises minimum detection limits and increases instrument downtime. One can eliminate the jet separator and shorten the analysis by using a narrow bore column and splitting the effluent from the trap before it enters the column, or by using cryogenics. Splitting the sample can reduce the amount of analyte entering the column by 90-95%, so detection limits often are no better than with a jet separator, and a cryogenic system increases both cost and total cycle time.

A new approach to concentrating analytes is fully compatible with short, narrow bore capillary columns that yield sharper peaks and better minimum detection limits, and greatly reduce analysis time. Solid phase microextraction[®] (SPME), like purge and trap, is a solventless extraction procedure, but SPME does not require the complex instrumentation of purge and trap methodology. SPME simply involves immersing a phase-coated fused silica fiber into the liquid sample or the headspace above the sample, to adsorb the analytes of interest. The adsorbed analytes are thermally desorbed in the injection port of the GC and transferred to the capillary column.[▲] Selectivity can be altered by changing the phase type or thickness. For example, the small distribution constants and low polarity of chlorinated and aromatic VOCs dictate the use of a thick, nonpolar phase for efficient extraction (see the conditions in Table 1). Agitation, addition of salt, pH adjustment, and immersion of the fiber in the liquid sample improve recovery of difficult-to-extract compounds.

Figure A. Rapidly Screen a Sample for VOCs, Using SPME/Dual Column GC




Table 1. Linearity of Response Factors for US EPA Method 624 Volatile Compounds

No.	Compound	Column [▲]	Response Factors	
			Mean	% RSD
1.	Chloromethane	SPB-1	0.022	17.0
2.	Vinyl chloride	SPB-1	0.663	23.0
3.	Bromomethane	SPB-1	0.025	11.4
4.	Chloroethane	SPB-1	0.229	14.7
5.	Trichlorofluoromethane	SPB-1	0.022	8.3
6.	1,1-Dichloroethene	VOCOL	0.341	13.3
7.	Methylene chloride	VOCOL	0.040	14.7
8.	trans-1,2-Dichloroethene	VOCOL	0.354	15.3
9.	1,1-Dichloroethane	VOCOL	0.272	9.1
10.	Chloroform	VOCOL	0.106	12.1
11.	1,1,1-Trichloroethane	SPB-1	0.374	5.1
12.	Carbon tetrachloride	VOCOL	0.080	11.9
13.	1,2-Dichloroethane	SPB-1	0.183	7.8
14.	Benzene	SPB-1	1.951	5.1
15.	Trichloroethene	VOCOL	0.336	3.9
16.	1,2-Dichloropropane	VOCOL	0.529	3.4
17.	Bromodichloromethane	VOCOL	0.072	9.9
18.	2-Chloroethylvinylether	VOCOL	0.324	6.0
19.	cis-1,3-Dichloropropene	VOCOL	0.551	3.6
20.	Toluene	VOCOL	2.091	5.2
21.	trans-1,3-Dichloropropene	VOCOL	0.501	4.3
22.	1,1,2-Trichloroethane	VOCOL	0.247	3.4
23.	Tetrachloroethene	VOCOL	0.251	13.0
24.	Dibromochloromethane	VOCOL	0.060	6.1
25.	Chlorobenzene	SPB-1	1.543	6.5
26.	Ethylbenzene	SPB-1	1.892	14.0
27.	Bromoform	SPB-1	0.086	6.4
IS	1,4-Dichlorobutane (int. std.)			
28.	1,1,2,2-Tetrachloroethane	VOCOL	0.274	4.9
29.	1,3-Dichlorobenzene	VOCOL	1.021	16.9
30.	1,4-Dichlorobenzene	VOCOL	1.078	16.3
31.	1,2-Dichlorobenzene	VOCOL	1.032	17.4

▲Column used to quantify the analyte

Sample: US EPA 624 VOCs in 1.8mL saturated salt water (2mL vial) 25ppb-1ppm, 7 concentration points
 Fiber Type: 100µm polydimethylsiloxane (PDMS)
 Extraction: direct immersion of fiber in sample (5 min, rapid stirring)

For screening VOCs with nonspecific detectors, such as FIDs and TCDs, a dual column analysis on, for example, a nonpolar column and an intermediate polarity column provides better identification and quantification of the analytes. A 10m x 0.2mm x 1.2µm SPB™-1 column and a 10m x 0.2mm x 1.2µm VOCOL™ column each provided good resolution of US EPA Method 624 analytes without long analysis times. Figure A shows the dual column analysis of the Method 624 analytes at 50ppb, following extraction by SPME. All analytes were identified and quantified in only 6 minutes. The cool down time – about 4 minutes – was greatly shortened by starting the analysis at 40°C instead of at 35°C. The 10-minute cycle time for the analysis is compatible with the sample preparation time by SPME – 5 minutes for extraction and 3 minutes for desorption. SPME is equally compatible with the conditions required by a GC/MS system.

Because wastewater samples can contain analytes at concentrations ranging from trace ppb to ppm levels, a sample screening technique must be suitable for quantifying analytes over a wide range of concentrations. In a purge and trap instrument, analytes

at concentrations greater than 200ppb can saturate the trap and contaminate the valves and lines, requiring downtime to clean the system. SPME is effective over a wide range of analyte concentrations, and thus is well suited for screening samples on site or prior to purge and trap/GC/MS analysis. Samples found to be highly concentrated can be diluted prior to the formal analysis.

We determined average response factors for 31 volatile analytes in US EPA Method 624 at a concentration range of 25ppb-1ppm, using SPME to extract the analytes. Data are summarized in Table 1. The low percent relative standard deviations (% RSD) for most analytes indicate good linearity for the response factors for this range of concentrations. The % RSD for vinyl chloride is unusually high because vinyl chloride coelutes with methanol, the solvent used with the standard. Responses for vinyl chloride are more linear with specific detectors, such as a mass spectrometer or ELCD.

These results show that SPME is fast, easy, and compatible with short, narrow bore columns that provide fast analysis times. Volatile organic compounds can be extracted with good accuracy over a wide concentration range. Because the apparatus is portable and easy to use, SPME can be employed in the field for quick turnaround methods, or for screening a sample prior to GC/MS analysis. Precision and accuracy also make SPME effective in quantitative analyses. If your analyses involve these or similar analytes, SPME might be the ideal answer to your sample preparation needs.

Ordering Information:

Description	Cat. No.
SPME Holder	
Initially you must order both holder and fiber assembly. Holder is reusable indefinitely.	
For manual sampling	57330-U
For Varian 8100/8200 AutoSampler (requires Varian SPME upgrade kit)	57331
SPME Fiber Assembly (pk. of 3)	
100µm polydimethylsiloxane coating	
For manual sampling	57300-U
For Varian 8100/8200 AutoSampler	57301
Capillary GC Columns	
10m x 0.20mm ID, 1.2µm film	
SPB-1	24134-U
VOCOL	24129-U

■ Solid phase microextraction technology is licensed exclusively to Supelco (US patent pending; European patent #0523092).

Fused silica columns manufactured under HP US Pat. No. 4,293,415.

SPB and VOCOL are trademarks of Sigma-Aldrich Co.

▲Gaseous VOCs must be rapidly desorbed and introduced into the column if they are to be resolved without cryogenics. The injection port is a major source of band broadening that can be reduced by reducing the inlet volume. When we replaced the 2mm ID splitless liner in the standard splitless/split injection port in our GC/MS system with a 0.75mm ID liner, peak shape and height for the gaseous VOCs were greatly improved.

Notes 56

For more information, or current prices, contact your nearest Supelco subsidiary listed below. To obtain further contact information, visit our website (www.sigma-aldrich.com), see the Supelco catalog, or contact Supelco, Bellefonte, PA 16823-0048 USA.

ARGENTINA · Sigma-Aldrich de Argentina, S.A. · Buenos Aires 1119 AUSTRALIA · Sigma-Aldrich Pty. Ltd. · Castle Hill NSW 2154 AUSTRIA · Sigma-Aldrich Handels GmbH · A-1110 Wien
 BELGIUM · Sigma-Aldrich N.V./S.A. · B-2880 Bornem BRAZIL · Sigma-Aldrich Quimica Brasil Ltda. · 01239-010 São Paulo, SP CANADA · Sigma-Aldrich Canada, Ltd. · 2149 Winston Park Dr., Oakville, ON L6H 6J8
 CZECH REPUBLIC · Sigma-Aldrich s.r.o. · 186 00 Praha 8 DENMARK · Sigma-Aldrich Denmark A/S · DK-2665 Vallensbaek Strand FINLAND · Sigma-Aldrich Finland/YA-Kemia Oy · FIN-00700 Helsinki
 FRANCE · Sigma-Aldrich Chimie · 38297 Saint-Quentin-Fallavier Cedex GERMANY · Sigma-Aldrich Chemie GmbH · D-82041 Deisenhofen GREECE · Sigma-Aldrich (o.m.) Ltd. · Ilioupoli 16346, Athens
 HUNGARY · Sigma-Aldrich Kft. · H-1067 Budapest INDIA · Sigma-Aldrich Co. · Bangalore 560 048 IRELAND · Sigma-Aldrich Ireland Ltd. · Dublin 24 ISRAEL · Sigma Israel Chemicals Ltd. · Rehovot 76100
 ITALY · Sigma-Aldrich s.r.l. · 20151 Milano JAPAN · Sigma-Aldrich Japan K.K. · Chuo-ku, Tokyo 103 KOREA · Sigma-Aldrich Korea · Seoul MALAYSIA · Sigma-Aldrich (M) Sdn. Bhd. · Selangor
 MEXICO · Sigma-Aldrich Quimica S.A. de C.V. · 50200 Toluca NETHERLANDS · Sigma-Aldrich Chemie BV · 3330 AA Zwijndrecht NORWAY · Sigma-Aldrich Norway · Torshov · N-0401 Oslo
 POLAND · Sigma-Aldrich Sp. z o.o. · 61-663 Poznań PORTUGAL · Sigma-Aldrich Quimica, S.A. · Sintra 2710 RUSSIA · Sigma-Aldrich Russia · Moscow 103062 SINGAPORE · Sigma-Aldrich Pte. Ltd.
 SOUTH AFRICA · Sigma-Aldrich (pty) Ltd. · Jet Park 1459 SPAIN · Sigma-Aldrich Quimica, S.A. · 28100 Alcobendas, Madrid SWEDEN · Sigma-Aldrich Sweden AB · 135 70 Stockholm
 SWITZERLAND · Supelco · CH-9471 Buchs UNITED KINGDOM · Sigma-Aldrich Company Ltd. · Poole, Dorset BH12 4QH
 UNITED STATES · Supelco · Supelco Park · Bellefonte, PA 16823-0048 · Phone 800-247-6628 or 814-359-3441 · Fax 800-447-3044 · email:supelco@sial.com

H

Supelco is a member of the Sigma-Aldrich family. Supelco products are sold through Sigma-Aldrich, Inc. Sigma-Aldrich warrants that its products conform to the information contained in this and other Sigma-Aldrich publications. Purchaser must determine the suitability of the product for a particular use. Additional terms and conditions may apply. Please see the reverse side of the invoice or packing slip.