

## Development of a High Sensitivity, Sub-Picogram SPE LC-MS Method for Quantification of Mometasone from Human Plasma

Taposh G, PMN Rajesh, and Dr. Padmakar Wagh  
India Technology Centre, Waters India pvt. Ltd, Bangalore

### APPLICATION BENEFITS

- Use of UPLC™ columns for excellent resolution for mometasone furoate from endogenous interferences
- Fast, simple, and selective sample preparation using Oasis™ HLB Cartridges
- Linear, accurate, and precise quantitative performance
- High analytical sensitivity detection, achieving 0.5 pg/mL, using the Xevo™ TQ-XS equipped with UniSpray™ ion source

### WATERS SOLUTIONS

[ACQUITY™ UPLC I-Class FTN System](#)

[Xevo TQ-XS Mass Spectrometer](#)

[ACQUITY UPLC BEH Phenyl Column](#)

[Oasis HLB Cartridges](#)

[Waters™ Total Recovery Vials](#)

### KEYWORDS

Mometasone furoate, plasma, solid phase extraction, SPE, Oasis, quantitation, corticosteroid, bioanalysis, Xevo TQ-XS, UPLC, UniSpray

### INTRODUCTION

Mometasone furoate (Figure 1)<sup>1</sup> is a synthetic corticosteroid with the chemical name 9, 21-dichloro-11(Beta),17-dihydroxy-16(alpha)-methylpregna1,4-diene-3,20-dione 17-(2-furoate). It is a potent beta 2-agonist with anti-inflammatory activity, used in the treatment of asthma and/or chronic obstructive pulmonary disease.<sup>2</sup>




Figure 1. Mometasone furoate chemical structure.

With minimal bioavailability (<1 %) and very low circulating plasma concentrations (50 pg/mL) following a 100–400 µg inhaled dose, accurate quantification of mometasone from plasma can be challenging.<sup>3</sup> In this work, a robust, sensitive, and selective method was developed for the accurate quantification of mometasone furoate. This method uses UPLC separation, tandem quadrupole MS with UniSpray ionization, and selective solid-phase extraction (SPE) sample preparation, achieving a lower limit of quantitation (LLOQ) of 0.5 pg/mL from 600 µL of plasma.

## EXPERIMENTAL

### Sample preparation

Standards and quality control (QC) samples were prepared by spiking mometasone furoate (0.5 to 60 pg/mL) into commercially available human plasma. Calibration curve standards were prepared in duplicate to check the reproducibility, while five replicates were prepared for the QC's. 25 µL of mometasone furoate-d3 (5 ng/mL), which was used as internal standard (ISTD), was added to the plasma samples spiked with mometasone furoate (600 µL). Following internal standard addition, sample was treated with 200 µL methanol and mixed. The pretreated plasma sample was then extracted using the Waters Oasis HLB 1 cc Cartridges and the extraction protocol shown in Figure 2. Following extraction, samples were injected for LC-MS/MS analysis.




Figure 2. SPE extraction protocol for mometasone furoate from human plasma using Oasis HLB 1 cc Cartridges.

### LC-MS conditions

|                          |   |
|--------------------------|---|
| LC system:               | ACQUITY UPLC I-Class                                  |
| Column:                  | ACQUITY BEH Phenyl 130Å,<br>1.7 µm, 2.1 mm × 100 mm   |
| Column temp.:            | 50 °C   |
| Sample temp.:            | 10 °C   |
| Injection volume:        | 10 µL   |
| Mobile phase A:          | 0.1% Formic acid in<br>5 mm ammonium formate in water |
| Mobile phase B:          | Methanol  |
| Reconstitution solution: | 50:50 Methanol:water                                  |

### Gradient table

| Time (min) | Flow (mL/min) | Mobile phase A | Mobile phase B | Curve |
|------------|---------------|----------------|----------------|-------|
| 0.0        | 0.4           | 95             | 5              | 6     |
| 0.5        | 0.4           | 95             | 5              | 6     |
| 6.0        | 0.4           | 10             | 90             | 6     |
| 7.0        | 0.4           | 10             | 90             | 6     |
| 7.1        | 0.4           | 95             | 99             | 6     |
| 8.0        | 0.4           | 95             | 10             | 6     |

Table 1. LC gradient for mometasone furoate analysis.

### MS conditions

|                          |                               |
|--------------------------|-------------------------------|
| Mass spectrometer:       | Xevo TQ-XS Tandem Quadrupole  |
| Ionization:              | ESI+ with UniSpray ion source |
| Capillary voltage:       | 0.5 KV (+)                    |
| Desolvation temp.:       | 650 °C                        |
| Cone gas flow:           | 150 L/h                       |
| Desolvation gas flow:    | 1000 L/h                      |
| Collision cell pressure: | 3.8 × e-3 mbar                |

| Analyte                      | Precursor ion (m/z) | Product ion (m/z) | Cone voltage | Collision energy (eV) |
|------------------------------|---------------------|-------------------|--------------|-----------------------|
| Mometasone furoate           | 521.16              | 355.05            | 37           | 13                    |
| Mometasone furoate-d3 (ISTD) | 524.08              | 355.05            | 37           | 13                    |

Table 2. MRM transitions for mometasone and mometasone furoate-d3 (ISTD).

### Data management

|                          |             |
|--------------------------|-------------|
| Chromatography software: | MassLynx™   |
| Quantification software: | TargetLynx™ |

## RESULTS

In this work, we have developed a complete sample preparation and UPLC LC-MS/MS method for sensitive and accurate quantification of mometasone furoate from plasma. SPE extraction of mometasone furoate from plasma was achieved using Oasis HLB 1 cc Cartridges using the SPE extraction procedure shown in Figure 2. Oasis HLB is a polymeric reversed-phased sorbent that provides high capacity and added specificity to this method. Recovery for mometasone furoate using this SPE extraction method was ~85%. LC-MS/MS quantification was performed using a Waters™ Xevo TQ-XS tandem/triple quadrupole mass spectrometer, coupled to an ACQUITY UPLC I-Class System. Chromatographic separation was achieved with an ACQUITY UPLC BEH Phenyl 1.7  $\mu\text{m}$  Column, at a flow rate of 0.4 mL/min using a linear gradient (Table 1) with 5 mM ammonium formate in water containing 0.1% formic acid and methanol mobile phases. MRM transitions used for quantification are summarized in Table 2.

The Xevo TQ-XS Mass Spectrometer, equipped with a step-wave ion guide and UniSpray ion source, enabled improved ion sampling in the source, better ion transfer efficiency, and improved ionization. Use of a low dispersion UPLC I-Class System and sub-2- $\mu\text{m}$  UPLC column, coupled to the Xevo TQ-XS, afforded excellent resolution from endogenous matrix components, enhancing selectivity, and sensitivity. This sensitivity and selectivity is illustrated in Figure 3 for the LLOQ, 0.5 pg/mL extracted plasma sample, achieving a signal-to-noise (S/N) ratio of 19.7.




Figure 3. Representative mometasone furoate LLOQ (0.5 pg/mL) chromatogram, extracted from plasma, demonstrating excellent sensitivity and selectivity with S/N ratio of 19.7.

Using this SPE LC-MS method, quantitative performance was excellent. Calibration curves were linear ( $r^2 > 0.9959$ ) from 0.5–60 pg/mL with accuracies between 85–115% and CVs <15% for all points on the curve. Figure 4 illustrates this performance. At the same time, QC statistics easily met recommended bioanalytical method development guidelines,<sup>4</sup> with average precision and accuracy values <15%. This QC performance is highlighted in Table 3 for precision and accuracy (PA) batches, while chromatographic performance is illustrated in Figure 5.




Figure 4. Representative mometasone furoate calibration curve (0.5 to 60 pg/mL) extracted from plasma.




Figure 5. Representative QC's chromatograms for mometasone furoate extract from plasma, highlighting sensitivity and specificity.

|                 | Sample #           | LQC<br>(1.25 pg/mL) | MQC<br>(30 pg/mL) | HQC<br>(54 pg/mL) |
|-----------------|--------------------|---------------------|-------------------|-------------------|
| Batch 1         | 1                  | 1.26                | 29.82             | 51.26             |
|                 | 2                  | 1.31                | 30.06             | 50.92             |
|                 | 3                  | 1.31                | 29.62             | 50.71             |
|                 | 4                  | 1.62                | 31.16             | 51.03             |
|                 | 5                  | 1.23                | 30.21             | 50.46             |
|                 | Mean conc. (pg/mL) | 1.35                | 30.17             | 50.88             |
|                 | STD                | 0.157               | 0.596             | 0.306             |
|                 | % CV               | 11.66               | 1.97              | 0.60              |
| Mean % accuracy | 107.68             | 100.58              | 94.21             |                   |
| Batch 2         | 6                  | 1.25                | 29.08             | 48.99             |
|                 | 7                  | 1.34                | 28.74             | 48.77             |
|                 | 8                  | 1.23                | 29.25             | 44.48             |
|                 | 9                  | 1.56                | 29.52             | 50.17             |
|                 | 10                 | 1.15                | 29.83             | 48.82             |
|                 | Mean conc. (pg/mL) | 1.31                | 29.28             | 48.25             |
|                 | Std                | 0.157               | 0.416             | 2.182             |
|                 | % CV               | 12.04               | 1.42              | 4.52              |
|                 | Mean % accuracy    | 104.48              | 97.61             | 89.34             |

Table 3. Summary of quality control results for mometasone furoate extracted from human plasma for PA batches 1 and 2.

## CONCLUSION

This application highlights the development of a highly sensitive and robust LC-MS assay for the quantification of mometasone furoate extracted from plasma. The method described herein achieves an LLOQ of 0.5 pg/mL with a linear dynamic range of 0.5–60 pg/mL. The high sensitivity and linearity of method was attributed to extraction specificity of Oasis HLB solid phase extraction sample preparation, high resolution UPLC chromatographic separation with the ACQUITY UPLC I-Class System using a sub-2- $\mu$ m UPLC column, and sensitivity of the Xevo TQ-XS Mass Spectrometer with UniSpray Technology. This developed method has demonstrated its fit-for-purpose use in support of drug discovery and research.

## References

- O'Neil MJ (2006) *The Merck Index: An Encyclopedia of Chemicals, Drugs, and Biologicals*. (14th edn), Merck & Co. Inc, 4244.
- Sweetman SC (2007) *Martindale the Complete Drug Reference*. (35th edn), Pharmaceutical press, London, 925.
- Daley-Yates, PT (2015) Inhaled Corticosteroids: Potency, Dose Equivalence and Therapeutic Index. *Br J Clin Pharmacol*, 80: 372–380. doi: 10.1111/bcp.12637.
- Validated Method Development for Estimation of Formoterol Fumarate and Mometasone Furoate in Metered Dose Inhalation Form by High Performance Liquid Chromatography. Srinivasarao, Katari 1,2\*; Gorule, Vinayk3; Reddiah, Venkata Ch3, and Krishna, Venkata A3.

# Waters

THE SCIENCE OF WHAT'S POSSIBLE.™

Waters, The Science of What's Possible, ACQUITY, UPLC, Oasis, Unispray, MassLynx, TargetLynx, and Xevo are trademarks of Waters Corporation. All other trademarks are the property of their respective owners.

©2018 Waters Corporation. Produced in the U.S.A. September 2018 720006399EN AG-PDF

Waters Corporation  
34 Maple Street  
Milford, MA 01757 U.S.A.  
T: 1 508 478 2000  
F: 1 508 872 1990  
www.waters.com