

Instrument: FP828

Determination of Nitrogen/Protein in Meat and Plant-Based Meat Alternatives

LECO Corporation; Saint Joseph, Michigan USA

Introduction

Protein is one of the most significant nutrient components in food products. The accurate and precise determination of protein not only plays a role in the characterization of nutritional or dietary value in food materials but is also the key to the economic value of these materials. Protein in meat products is most commonly calculated using the measured nitrogen content in the sample and a multiplier or conversion factor (commonly 6.25). Nitrogen determination is performed using either the classical wet chemical method (Kjeldahl) or a combustion method (Dumas). The Kjeldahl method involves sample digestion, distillation, and ammonia determination typically by titration. This method involves time-consuming sample preparation and the use of hazardous materials. The LECO FP828 is a nitrogen determinator that utilizes an automated combustion (Dumas) method and provides accurate and precise results in approximately three minutes. This eliminates involved sample preparation and the use of hazardous materials resulting in a rapid, cost-effective method for the quality control of the production of meat and plant-based protein products.

Instrument Model and Configuration

The LECO FP828 is a combustion Nitrogen/Protein determinator that utilizes a pure Oxygen environment in a vertical quartz furnace, ensuring complete combustion and superior analyte recovery. A thermoelectric cooler removes moisture from the combustion gases before they are collected in a ballast. The combustion gases equilibrate and mix in the ballast before a representative aliquot (3 cm³ or 10 cm³ volume) of the gas is extracted and introduced into a flowing stream of inert gas (Helium or Argon) for analysis. The aliquot gas is carried to a thermal conductivity cell (TC) for the detection of Nitrogen (N₂).

Thermal conductivity detectors work by detecting changes in the thermal conductivity of the analyte gas compared to a reference/carrier gas. The greater the difference between the thermal conductivity of the carrier gas and the analyte gas, the greater sensitivity of the detector.

The FP828 supports either the use of Helium or Argon as the instrument's carrier gas. When used as a carrier gas, Helium provides the highest sensitivity, and the best performance at the lower limit of the Nitrogen range.

The thermal conductivity difference between Argon and Nitrogen is not as great as the thermal conductivity difference between Helium and Nitrogen, therefore the detector is inherently less sensitive when using Argon as a carrier gas.

The FP828 offers the additional advantage of utilizing either a 10 cm³ aliquot dose loop or a 3 cm³ aliquot dose loop within the instrument's gas collection and handling system. The 10 cm³ aliquot dose loop optimizes the system for the lowest Nitrogen range and provides the best precision. The 3 cm³ aliquot dose loop extends reagent life expectancy by approximately three-fold when compared to the 10 cm³ aliquot dose loop, while providing the lowest cost-per-analysis.

Note: When changing carrier gas type, refer to the 828 Series Operator's Instruction Manual for the procedure on setting the gas flow rate. When using the FP828 Performance model, the aliquot dose loop size is changed by selecting the desired aliquot dose loop size in the software's Method Parameters. When using the FP828 Base model, the desired dose loop is installed by the operator.

Method Reference*

AOAC 992.15 – Crude Protein in Meat and Meat Products Including Pet Foods (Combustion Method)*

*A modified version of the AOAC 992.15 official method was utilized to generate this application note.

Sample Preparation

Samples must be of a uniform consistency to produce suitable results. Meat samples should be prepared in accordance with AOAC 983.18 Method: Meat and Meat Products - Preparation of Samples. Reference materials should be prepared as directed by the certificate, prior to analysis.

Accessories

502-186 Small Tin Foil Cups, 501-614 Spatula

Reference Materials

LCRM[®], LRM[®], NIST, or other suitable reference materials.

General Parameters**

Gas Type	Helium or Argon	
Furnace Temperature	950 °C	
Afterburner Temperature	850 °C	
Nominal Mass	1.0000 g	
Purge Cycles	Helium	Argon
Ballast Equilibrate Time	10 s	10 s
Ballast Not Filled Timeout	300 s	300 s
Aliquot Loop Fill Pressure Drop	175 mm Hg	200 mm Hg
Aliquot Loop Equilibrate Time	4 s	6 s
Dose Loop Size	10 cm ³ or 3 cm ³	10 cm ³ or 3 cm ³
Interleave Analysis	Yes	Yes
Sample Drop Detection	Disabled	Disabled

**Refer to 828 Series Operator's Instruction Manual for Parameter definitions.

Element Parameters**

	Helium	Argon
Integration Delay	3 s	5 s
Starting Baseline	15 s	15 s
Post Baseline Delay	15 s	15 s
Use Comparator	No	No
Integration Time	45 s	65 s
Use Endline	Yes	Yes
Endline Delay	25 s	25 s
Ending Baseline	15 s	15 s
Use Profile Blank	---	Yes

**Refer to 828 Series Operator's Instruction Manual for Parameter definitions.

Burn Profile

Performance Model

Burn Step	Furnace Flow	Time
1	5.00 L/min	30 s
2	1.00 L/min	30 s
3	5.00 L/min	End

Base Model

Burn Step	Furnace Flow	Time
1	High	30 s
2	Medium	30 s
3	High	End

Procedure

1. Prepare instrument for operation as outlined in the operator's instruction manual.
2. Condition the System.
 - a. Select three to five replicates in the Login screen.
 - b. Weigh an appropriate mass of a reference material or sample material into a 502-186 Tin Foil Cup and enter the sample mass and identification into the Login screen.
 - c. Seal the foil cup in a manner to minimize entrapped atmosphere by twisting the top edges of the foil together and transfer it to the appropriate position in the sample carousel.
 - d. Perform steps 2b through 2c for each conditioning sample to be analyzed.
 - e. Initiate the analysis sequence.
3. Determine Blank.
 - a. Select five or more blank replicates in the Login screen.
 - b. Initiate the analysis sequence.
 - c. Set the blank using at least five blank results following the procedure outlined in the operator's instruction manual.

Note: The standard deviation of the last five blanks should be less than or equal to 0.001% (10 ppm) when utilizing Helium as a carrier gas, and less than or equal to 0.005% (50 ppm) when utilizing Argon as a carrier gas. Additional blanks beyond the recommended five may be required in order to achieve the recommended precision.

4. Calibrate/Drift Correct.

- a. Select the desired number of calibration/drift replicates in the Login screen (minimum of five).
- b. Weigh an appropriate mass of a suitable reference material into a 502-186 Tin Foil Cup and enter the sample mass and identification into the Login screen.
- c. Seal the foil cup in a manner to minimize entrapped atmosphere by twisting the top edges of the foil together and transfer it to the appropriate position in the sample carousel.
- d. Perform steps 4b through 4c a minimum of five times.
- e. Initiate the analysis sequence.
- f. Calibrate or Drift Correct the instrument following the procedure outlined in the operator's instruction manual.
- g. Verify the calibration or drift correction by analyzing an appropriate mass of another/different suitable reference material and confirm that the results are within the acceptable tolerance range.

5. Analyze Samples.
 - a. Select the desired number of sample replicates in the Login screen.
 - b. Weigh ~0.25 g of a sample into a 502-186 Tin Foil Cup and enter sample mass and identification into the Login screen.
 - c. Seal the foil cup in a manner to minimize entrapped atmosphere by twisting the top edges of the foil together and transfer it to the appropriate position in the sample carousel.
 - d. Perform steps 5b through 5c for each sample to be analyzed.
 - e. Initiate the analysis sequence.

TYPICAL RESULTS

Data was generated utilizing a linear, force through origin calibration using ~0.25 g of LECO 502-896 (Lot 1007) LCRM EDTA (9.59% N). The calibration was verified using ~0.1 g of LECO 502-642 (Lot 1020) LCRM Phenylalanine (8.46% N). A protein factor of 6.25[†] was used for all samples to calculate the protein content.

	10 cm ³ Helium			3 cm ³ Helium			10 cm ³ Argon			3 cm ³ Argon		
	Mass (g)	% N	% Protein	Mass (g)	% N	% Protein	Mass (g)	% N	% Protein	Mass (g)	% N	% Protein
Ham Deli Meat	0.2774	2.98	18.6	0.2672	2.93	18.3	0.2669	2.92	18.3	0.2445	2.95	18.5
	0.2679	2.96	18.5	0.2764	2.98	18.6	0.2502	2.88	17.9	0.2418	3.02	18.9
	0.2654	2.92	18.3	0.2884	2.95	18.4	0.2510	2.93	18.4	0.2417	2.98	18.6
	0.2534	2.98	18.6	0.2611	2.91	18.2	0.2686	2.96	18.5	0.2684	2.96	18.5
	0.2591	2.92	18.3	0.2628	3.00	18.7	0.2877	2.98	18.8	0.2462	2.92	18.2
	Avg=	2.95	18.5	Avg=	2.95	18.5	Avg=	2.94	18.4	Avg=	2.96	18.5
	<i>s=</i>	0.03	0.2	<i>s=</i>	0.04	0.2	<i>s=</i>	0.04	0.3	<i>s=</i>	0.04	0.2
Corned Beef Deli Meat	0.2693	3.33	20.8	0.2795	3.39	21.2	0.2564	3.35	21.0	0.2683	3.30	20.6
	0.2702	3.31	20.7	0.2592	3.37	21.0	0.2565	3.31	20.7	0.2693	3.38	21.1
	0.2790	3.32	20.8	0.2727	3.32	20.8	0.2690	3.35	21.0	0.2684	3.36	21.0
	0.2618	3.31	20.7	0.2743	3.34	20.9	0.2501	3.34	20.9	0.2738	3.38	21.1
	0.2720	3.22	20.1	0.2595	3.41	21.3	0.2611	3.41	21.3	0.2770	3.39	21.2
	Avg=	3.30	20.6	Avg=	3.37	21.0	Avg=	3.35	21.0	Avg=	3.36	21.0
	<i>s=</i>	0.04	0.3	<i>s=</i>	0.04	0.2	<i>s=</i>	0.04	0.2	<i>s=</i>	0.04	0.2
Turkey Deli Meat	0.2520	3.22	20.1	0.2842	3.17	19.8	0.2510	3.23	20.2	0.2536	3.29	20.6
	0.2620	3.21	20.0	0.2595	3.20	20.0	0.2731	3.32	20.8	0.2667	3.25	20.3
	0.2641	3.24	20.2	0.2663	3.25	20.3	0.2731	3.24	20.3	0.2620	3.23	20.2
	0.2799	3.28	20.5	0.2559	3.29	20.6	0.2580	3.27	20.4	0.2591	3.22	20.1
	0.2698	3.21	20.1	0.2747	3.28	20.5	0.2538	3.21	20.1	0.2580	3.22	20.1
	Avg=	3.23	20.2	Avg=	3.24	20.2	Avg=	3.25	20.3	Avg=	3.24	20.3
	<i>s=</i>	0.03	0.2	<i>s=</i>	0.05	0.3	<i>s=</i>	0.04	0.3	<i>s=</i>	0.03	0.2
Roast Beef Deli Meat	0.2781	3.57	22.3	0.2637	3.64	22.8	0.2796	3.60	22.5	0.2592	3.59	22.5
	0.2899	3.61	22.6	0.2726	3.66	22.9	0.2616	3.58	22.4	0.2842	3.54	22.2
	0.2761	3.59	22.5	0.2699	3.58	22.4	0.2459	3.54	22.1	0.2691	3.62	22.6
	0.2966	3.59	22.4	0.2662	3.61	22.5	0.2591	3.60	22.5	0.2645	3.57	22.3
	0.2616	3.58	22.4	0.2689	3.63	22.7	0.2841	3.56	22.3	0.2817	3.60	22.5
	Avg=	3.59	22.4	Avg=	3.62	22.7	Avg=	3.58	22.4	Avg=	3.59	22.4
	<i>s=</i>	0.01	0.1	<i>s=</i>	0.03	0.2	<i>s=</i>	0.02	0.2	<i>s=</i>	0.03	0.2
Veggie Burger	0.2742	2.87	17.9	0.2980	2.91	18.2	0.2418	2.80	17.5	0.2925	2.94	18.4
	0.2746	2.90	18.1	0.2626	2.85	17.8	0.2905	2.95	18.4	0.2584	2.87	17.9
	0.3084	2.95	18.4	0.2999	2.99	18.7	0.2743	2.90	18.1	0.2898	2.94	18.3
	0.2510	2.90	18.1	0.2577	2.91	18.2	0.2767	2.93	18.3	0.2859	2.89	18.1
	0.2866	2.83	17.7	0.2998	2.84	17.7	0.2951	2.99	18.7	0.2477	2.90	18.1
	Avg=	2.89	18.1	Avg=	2.90	18.1	Avg=	2.91	18.2	Avg=	2.91	18.2
	<i>s=</i>	0.04	0.3	<i>s=</i>	0.06	0.4	<i>s=</i>	0.07	0.5	<i>s=</i>	0.03	0.2

[†]Protein factor was obtained from the United States Department of Agriculture, Circular No. 183. The choice of protein factor to be used for determining protein content in different materials is the subject of some debate. As a result, if being used for commerce, the value of this conversion factor should be part of the contractual agreement between buyer and seller.



LECO Corporation | 3000 Lakeview Avenue | St. Joseph, MI 49085 | Phone: 800-292-6141 | 269-985-5496

info@leco.com • www.leco.com | ISO-9001:2015 Certified | LECO is a registered trademark of LECO Corporation.

LCRM = LECO Certified Reference Material; LRM = LECO Reference Material and are registered trademarks of LECO Corporation.