

Application News

Inductively Coupled Plasma Mass Spectrometer

Determination of Additive and Trace Elements in Lubricating Oil Using ICP-MS

Kosuke Naka

User Benefits

- ◆ Metal elements in lubricating oil can be accurately analyzed by dilution with organic solvents.
- ◆ Solvent-resistant peristaltic pump tubing enables the addition of internal standard elements in-line, even for organic solvent analysis, thereby eliminating the need for manual addition.

■ Introduction

Organic metal compounds are added to lubricating oil to provide various functional properties. Therefore, monitoring the concentrations of these additives is important for ensuring quality control. In addition, the analysis of wear metals and contaminant metals in lubricating oils (engine oils) used in automobiles and ships is a crucial method for assessing the condition of the engine. Traditionally, these analyses have utilized ICP Optical Emission Spectrometry (ICP-OES) as outlined in ASTM D4951-14¹⁾ and ASTM D5185-18²⁾. However, for the analysis of trace metal elements that are difficult to detect with ICP-OES, the more sensitive ICP mass spectrometry (ICP-MS) is more suitable.

In this Application News, lubricating oil was diluted with organic solvents, and the additives and trace elements in the oil were analyzed using the ICP-MS 2050 (Fig. 1). To verify the accuracy of the analytical results, spike recoveries were also evaluated.



Fig. 1 ICPMS-2050 and AS-20

■ Samples and Reagents

• Samples

Commercially available automotive engine oil (mineral oil) was used.

• Diluent

PremiSolv (CONOSTAN) was used.

• Reagents for standards

The oil-based mixed standard solution S-21 and the oil-based single-element standard solutions for Li, P, K, Co, Sb, Be, Y, In, and Bi (CONOSTAN), and the oil-based single-element standard solution for S, Sc (LGC Standards) were used.

■ Preparation of Samples and Standards

• Sample Preparation (for Trace Element Analysis)

Samples were diluted 50-fold (w/w) with PremiSolv to prepare them for analysis of trace elements. Spiked samples were prepared similarly, with one solution containing trace elements excluding sulfur (S), and another solution containing only S.

• Sample Preparation (for Additive Element Analysis)

The sample was diluted 5000-fold (w/w) with PremiSolv to prepare a sample solution for analysis of the additive elements.

• Calibration standards

S-21 and single-element standard solutions were mixed and diluted with PremiSolv to prepare the calibration standards. The concentrations of each element in the calibration standards are shown in Table 1.

• Internal standard solution

Oil-based single-element standard solutions of Be, Sc, Y, In, and Bi were mixed and diluted with PremiSolv. The concentrations of Be and Sc in the internal standard solution are 1000 µg/kg, while the concentrations of Y and In are 100 µg/kg, and Bi is 20 µg/kg.

■ Configurations and Analytical Conditions

Table 2 shows the configuration of the ICP-MS. The "Organic Solvent Injection System" was used as the sample introduction system. By using a platinum sampling cone, damage to the cone from prolonged exposure to organic solvents can be minimized. Additionally, the use of solvent-resistant peristaltic pump tubing enables the in-line addition of internal standard elements, thereby streamlining the sample preparation process.

Table 3 shows the analytical conditions. To prevent the precipitation of carbon from the organic solvent in the interface, a mixed gas of argon (70 %) and oxygen (30 %) was introduced into the quadrupole structure organic solvent torch. To reduce the effects of non-spectral interferences, such as ionization interference from highly concentrated additive elements in lubricating oil, dilution gas was introduced.

Table 1 Concentrations of Analytes in Calibration Standards

Elements	Calibration Standards (µg/kg)						
	STD0	STD1	STD2	STD3	STD4	STD5	STD6
Li, Na, Al, K, Ti, V, Cr, Mn, Fe, Co, Ni, Cu, Cd, Sn, Sb, Pb	0	10	50	200			
B, Mg, Si, Ca, Zn, Mo	0		50	200			
P	0	1,000	5,000	20,000			
S	0				10,000	50,000	100,000

Table 2 ICP-MS System Configuration

Instrument	: ICPMS-2050
Nebulizer	: Nebulizer DC04
Chamber	: Cyclone Chamber
Torch	: Organic Solvent Torch
Sampling Cone	: Platinum
Skimmer Cone	: Nickel
Auto Sampler	: AS-20 (Rinse station for organic solvents) I.D. 0.76 mm for sample ¹
Peristaltic Pump Tubing	: I.D. 0.38 mm for internal standards ² (materials: Solva PVC) Online Internal Standard Kit (for Organic Solvents) ³
Internal Standard Elements	: (sample: internal standard = about 4:1)

I.D.: Internal Diameter

¹*: P/N: S018-31558-61²*: P/N: S018-31558-62³*: P/N: S211-95010-42

Table 3 Analytical Conditions

RF Power	: 1.60 kW
Sampling Depth	: 8.0 mm
Plasma Gas Flowrate	: 20.0 L/min
Auxiliary Gas Flowrate	: 0.50 L/min
Carrier Gas Flowrate	: 0.40 L/min
Dilution Gas Flowrate	: 0.20 L/min
Ar-O ₂ Mixed Gas Flowrate	: 0.35 L/min
Chamber Temperature	: -5 °C
Pump Rotation Speed	: 15 r.p.m. (Low) / 45 r.p.m. (High)
Cell Condition	: No Gas
Cell Gas	: -
Cell Gas Flowrate (mL/min)	: 6.0
Cell Voltage (V)	: -35
Energy Filter (V)	: 7.0
	He
	H ₂ -A
	H ₂ -B

■ Detection Limit

The instrument detection limits (IDLs) are shown in Table 4. The IDLs were calculated as the concentration that gives a signal equivalent to three times the standard deviation (σ) of the calibration blank sample (STD0).

■ Quantitative Analysis and Spike Recoveries

The trace metal elements in engine oil were quantitatively analyzed using the calibration standards shown in Table 1. The quantitative results were converted to the concentrations in engine oil by multiplying the dilution factors (trace elements: 49.89 times, additive elements: 4960 times). Additionally, to confirm the accuracy of the measurement, spike recoveries were evaluated. Table 5 shows the results of the quantitative analysis and the spike recoveries.

Good recoveries, ranging from 88 % to 116 %, were obtained, confirming that the influence of the inorganic matrix derived from the additive elements was sufficiently minimized. These results indicate that trace metal elements in engine oil can be accurately analyzed using only dilution with organic solvents.

■ Conclusion

In this Application News, the analysis of metal elements in lubricating oil was performed using the ICPMS-2050 with the organic solvent injection system. Good spike recoveries were achieved, confirming that trace metal elements in lubricating oil can be accurately quantified with a simple sample preparation involving only dilution with organic solvents. Since it is possible to analyze without the complex procedures required for acid decomposition, the risk of volatilization or contamination of the analytes during the sample preparation process is reduced. Additionally, the use of solvent-resistant peristaltic pump tubing enables the in-line addition of internal standard elements, further reducing the time required for sample preparation.

Table 4 Detection Limits (DLs)

Elements	Cell Condition	Internal Standards	IDLs (µg/kg)
⁷ Li	No Gas	⁹ Be	0.1
¹⁰ B	No Gas	⁹ Be	0.4
²³ Na	No Gas	⁴⁵ Sc	0.7
²⁴ Mg	H ₂ -A	⁴⁵ Sc	1
²⁷ Al	He	⁴⁵ Sc	0.6
²⁸ Si	H ₂ -A	⁴⁵ Sc	4
³¹ P	He	⁴⁵ Sc	40
³⁴ S	H ₂ -B	⁴⁵ Sc	1,000
³⁹ K	H ₂ -A	⁴⁵ Sc	0.8
⁴⁴ Ca	H ₂ -A	⁴⁵ Sc	2
⁴⁹ Ti	He	⁴⁵ Sc	0.9
⁵¹ V	H ₂ -A	⁴⁵ Sc	0.03
⁵² Cr	H ₂ -A	⁴⁵ Sc	0.03
⁵⁵ Mn	H ₂ -A	⁴⁵ Sc	0.04
⁵⁶ Fe	H ₂ -A	⁴⁵ Sc	0.2
⁵⁹ Co	He	⁴⁵ Sc	0.03
⁶⁰ Ni	He	⁴⁵ Sc	0.3
⁶³ Cu [*]	He	⁴⁵ Sc	0.2
⁶⁶ Zn	He	⁴⁵ Sc	0.7
⁹⁵ Mo	He	⁸⁹ Y	0.03
¹¹¹ Cd	H ₂ -B	¹¹⁵ In	0.05
¹¹⁸ Sn	He	¹¹⁵ In	0.06
¹²¹ Sb	He	¹¹⁵ In	0.02
¹³⁸ Ba	He	¹¹⁵ In	0.03
²⁰⁸ Pb	He	²⁰⁹ Bi	0.05

IDL (Instrument Detection Limit): 3σ (standard deviation of STD0) \times slope of calibration curve

*: In case that a large amount of Zn is contained, Cu may be affected by the peak tailing from Zn. To avoid peak tailing, Cu is measured in high resolution mode with double the standard resolution.

<References>

- 1) ASTM D4951-14 Standard Test Method for Determination of Additive Elements in Lubricating Oils by Inductively Coupled Plasma Atomic Emission Spectrometry
- 2) ASTM D5185-18 Standard Test Method for Multielement Determination of Used and Unused Lubricating Oils and Base Oils by Inductively Coupled Plasma Atomic Emission Spectrometry (ICP-AES)

<Related Applications>

1. Determination of Elemental Impurities in Petroleum Distillates using ICP-MS ~ASTM D8110-17~ Application News No.01-00922
2. Analysis of Additive Elements in Lubricating Oil According to ASTM D4951: ICPE-9820 [Application News No.J111A](#)
3. Analysis of Additive Elements, Wear Metals, and Contaminants in Used Lubricating Oil According to ASTM D5185: ICPE-9820 [Application News No.J114A](#)

Table 5 Quantitative Results in Lubricating Oil and Spike Recoveries

	Elements	IDLs ($\mu\text{g}/\text{kg}$)	Spike Conc. ($\mu\text{g}/\text{kg}$)	Unspiked Sample ($\mu\text{g}/\text{kg}$)	Spiked Sample ($\mu\text{g}/\text{kg}$)	Spike Recoveries (%)	Results in Engine Oil (mg/kg)
Trace Elements + P, S (Analyzed in a 50-fold diluted sample)	⁷ Li	0.1	50	1.3	55.8	109	0.65
	²³ Na	0.7	50	23.8	71.7	96	1.19
	²⁷ Al	0.6	50	12.7	64.3	103	0.634
	²⁸ Si	4	50	110	168	116	5.49
	³¹ P	40	5,000	3,770	9,480	114	188
	³⁴ S	1,000	50,000	53,800	106,000	104	2,680
	³⁹ K	0.8	50	10.2	62.6	105	0.509
	⁴⁹ Ti	0.9	50	N.D.	50.6	101	N.D.
	⁵¹ V	0.03	50	0.60	47.0	93	0.030
	⁵² Cr	0.03	50	0.33	44.3	88	0.017
	⁵⁵ Mn	0.04	50	5.15	56.6	103	0.257
	⁵⁶ Fe	0.2	50	22.0	71.4	99	1.10
	⁵⁹ Co	0.03	50	0.05	49.0	98	0.003
	⁶⁰ Ni	0.3	50	1.5	51.6	100	0.075
	⁶³ Cu	0.2	50	0.3	47.9	95	0.02
	¹¹¹ Cd	0.05	50	0.09	53.9	108	0.005
	¹¹⁸ Sn	0.06	50	0.24	52.6	105	0.012
Concentrated Elements (Analyzed in a 5000-fold diluted sample)	¹²¹ Sb	0.02	50	0.27	53.5	106	0.014
	¹³⁸ Ba	0.03	50	0.24	50.5	101	0.012
	²⁰⁸ Pb	0.05	50	0.59	50.8	100	0.029
	¹⁰ B	0.4		38.9			193
	²⁴ Mg	1		68.5			340
	⁴⁴ Ca	2		199			987
	⁶⁶ Zn	0.7		144			714
	⁹⁵ Mo	0.03		43.5			216

N.D.: Not Detected

Spike Recovery (%): (spiked sample – unspiked sample) / spike concentration $\times 100$ 

Shimadzu Corporation

www.shimadzu.com/an/

For Research Use Only. Not for use in diagnostic procedures.

This publication may contain references to products that are not available in your country. Please contact us to check the availability of these products in your country.

The content of this publication shall not be reproduced, altered or sold for any commercial purpose without the written approval of Shimadzu. See <http://www.shimadzu.com/about/trademarks/index.html> for details.

Third party trademarks and trade names may be used in this publication to refer to either the entities or their products/services, whether or not they are used with trademark symbol "TM" or "®".

Shimadzu disclaims any proprietary interest in trademarks and trade names other than its own.

The information contained herein is provided to you "as is" without warranty of any kind including without limitation warranties as to its accuracy or completeness. Shimadzu does not assume any responsibility or liability for any damage, whether direct or indirect, relating to the use of this publication. This publication is based upon the information available to Shimadzu on or before the date of publication, and subject to change without notice.

01-00921-EN

First Edition: Aug. 2025

➤ Please fill out the survey

Related Products

Some products may be updated to newer models.



➤ ICPMS-2040 Series
/ ICPMS-2050 Series
Inductively Coupled Plasma Mass
Spectrometry



➤ LabSolutions™
ICPMS Ver. 2

Related Solutions

Hydrocarbon

➤ Processing Industry
(Petrochemical, Ch

➤ Petroleum refinery

➤ Automotive

➤ Price Inquiry

➤ Product Inquiry

➤ Technical Service /
Support Inquiry

➤ Other Inquiry