

Maintaining Your Agilent GC's Split/Splitless Injection Port



**What's going on in there anyway & how do I
stay in control?**

Presentation Expectations

- The #1 inlet used by far is the Split/Splitless inlet, so we will focus on this inlet but allude to other inlets as appropriate.
- Many inlet parts are “consumables,” they have to be replaced for optimal **System** and **Method** performance.
- Replacement frequency varies with selection and the samples injected
- Presentation offers set-up and maintenance “tips” to improve performance and inlet robustness.
- Will treat root causes of performance degradation individually

SAMPLE INJECTION


Goals

- Introduce sample into the column
- Reproducible
- No efficiency losses
- Representative of sample




Influence of Injection Efficiency

Short
Concentrated



Solute Bands




Long
Diffuse

Same column, same chromatographic conditions

6850/6890 Split Injection

Inlet pressure setting



SPLIT

Flow Considerations

- 1-3 mL/min into the column
- 10-200 mL/min in the inlet (mostly split vent)
- 0.3-6 sec sweep rate
- Results in high efficiency

SPLIT

Split Ratio


- Ratio of split vent and column gas flows
- Approximates the amount of sample splitting
- Split ratio calculations are inaccurate

LOW SPLIT RATIO

mm I.D.	Lowest ratio
0.10	1:50 - 1:75
0.18 - 0.25	1:10 - 1:20
0.32	1:8 - 1:15
0.53	1:2 - 1:5

HIGH SPLIT RATIO

- Limited by flow controller
- Range usually 1:200 to 1:1000



**Gold plated inlet seal with
cross, p/n 5182-9652**



SPLIT RATIO COMPARISON

	High Ratio	Low Ratio
Sample into column	Low	High
Efficiency	High	Low
Discrimination	High	Low
Carrier gas use	High	Low

SPLIT Injection Characteristics

- Little backflash
- Highly Efficient Sample Transfer
- Inlet Discrimination
- Discrimination affected by many variables

Inlet Liners – Volume Considerations

Glass Inlet Liners provide an “inert” space for liquid samples to be uniformly vaporized to a gas and moved to the column.

Liquid-gas phase change involves a significant change in volume.

Gaseous sample volume depends on

- Injection volume
- Solvent type
- Column head pressure
- Inlet temperature
- **These aspects should be optimized for your sample volume and application.**

Solvent (1 μ L, ambient)	Volume (μ L at 250°C and 20psig)
n-Hexane	140
Acetone	245
Acetonitrile	350
Methanol	450

See “A Practical Guide to the Care, Maintenance, and Troubleshooting of Capillary
Revised Edition, by Dean Rood, Wiley-VCH, New York, 2001.



SPLIT Liners


Discrimination

- Liner design influences discrimination
- Greater flow disruption = Less discrimination
- Greater thermal mass = Less discrimination

SPLIT LINER




Straight tube



Straight tube with glass wool




Fixed glass wool, low pressure drop



Inverted cup

SPLIT LINER Packed With Glass Wool




DB-1, 15 m x 0.25 mm I.D., 0.25 μ m

60°C for 1 min, 60-180°C at 20°/min; Helium at 30 cm/sec

1. n-heptane 2. toluene 3. n-decane 4. n-butylbenzene 5. n-tridecane 6. biphenyl

SPLIT LINER Inverted Cup




DB-1, 15 m x 0.25 mm I.D., 0.25 μ m

60°C for 1 min, 60-180°C at 20°/min; Helium at 30 cm/sec

1. n-heptane 2. toluene 3. n-decane 4. n-butylbenzene 5. n-tridecane 6. biphenyl

SPLIT LINER Straight Tube



DB-1, 15 m x 0.25 mm I.D., 0.25 μ m

60°C for 1 min, 60-180°C at 20°/min; Helium at 30 cm/sec

1. n-heptane 2. toluene 3. n-decane 4. n-butylbenzene 5. n-tridecane 6. biphenyl

SPLIT LINER

Peak Area Ratios


	Peak Area*
Glass Wool	0.76
Inverted Cup	0.82
Straight	0.75

**Peak #1 Area/Peak #6 Area*

SPLIT


Silylated Glass Wool

- Traps non-volatile materials and mixes sample
- Peak shape and discrimination affected by amount, location and packing density




INJECTOR DISCRIMINATION

nC_8 to nC_{40} Alkanes in CH_2Cl_2




Oven: 35°C for 4 min, 35-320°C at 10°/min, 320°C for 5 min
Carrier Gas: Helium at 5.0 mL/min

ADD 10mg GLASS WOOL TO THE INJECTION PORT LINER



Oven: 35C for 4 min, 35-320°C at 10°/min, 320°C for 5 min
Carrier Gas: Helium at 5.0 mL/min

LARGER PLUG OF GLASS WOOL IN THE LINER



Oven: 35°C for 4 min, 35-320°C at 10°/min, 320°C for 5 min
Carrier Gas: Helium at 9.5 mL/min

GLASS WOOL

Placement in Liner

Near top of liner:

- Wipes syringe needle of sample
- Can improve injector precision
- Helps to prevent backflash

Near bottom of liner:

- Helps in volatilization of high MW components
- Increases mixing

GLASS WOOL

Considerations

- Always use deactivated (silylated) wool
- Borosilicate or quartz material?



GLASS WOOL

Liner Packing Recommendations

- Amount, size and placement must be consistent for consistent results
- Can be broken upon installation into the liner, exposing active sites
- Liner deactivation with glass wool plug in place is ideal



SPLIT INJECTOR

Temperature

- Hot enough to rapidly vaporize the sample
- May degrade sample or result in injector contamination if too hot
- Typically 200-250°C

SPLIT INJECTOR

Temperature




DB-1, 15 m x 0.25 mm I.D., 0.25 μ m

60°C for 1 min, 60-180°C at 20°/min; Helium at 30 cm/sec

1. n-heptane 2. toluene 3. n-decane 4. n-butylbenzene 5. n-tridecane 6. biphenyl

SPLIT INJECTOR

Injection Volume



SPLIT INJECTOR

Injection Volume

- Injection volume is not linear
- Inject same volume for all samples and standards for accurate and precise results

SPLIT


Summary

- Concentrated samples
- High efficiency
- All columns
- Liners - be consistent



6850/90 Splitless Injection

Until the end of the splitless purge delay. . .



Splitless Injection Overview

- For trace level analysis.
- Use split/splitless injection port in the splitless mode (split vent closed).
- The dilute sample is injected, the sample is volatilized, and majority of analytes condense on column.
- Later, the split vent is opened and residual solvent is vented.
- Timing, carrier and split vent flows, and oven temperature program are important.
- Sample has longer residence time in the heated inlet giving more opportunity to vaporize high boiling sample components compared to split injection.

SPLITLESS

Carrier Gas Flow Considerations At Injection

- 1-2 mL/min in the inlet and into column
- 1-2 minutes sweep rate
- Long sample residence time
- Slow sweep rate = Poor efficiency







SPLITLESS INJECTOR

Purge Activation Time

- Purges injector of residual sample
- Reduces solvent front size
- Typically 0.25 - 1.5 minutes
- Not linear in amount injected/unit time


Longer purge time = More sample in column and larger solvent front

Splitless Injection Liners

Liner	Part No.	Comments
	<p>5181-3316</p>	<p>Single taper, deactivated, 900μL volume. Taper isolates sample from metal seal, reducing breakdown of compounds that are active with metals. For trace samples, general application.</p>
	<p>5062-3587</p>	<p>Single taper, deactivated, with glass wool, 900μL volume. Glass wool aides volatilization and protects column. For trace (dirty) samples.</p>
	<p>5181-3315</p>	<p>Double taper, deactivated, 800μL volume. Taper on inlet reduces chance for backflash into carrier gas lines. High efficiency liner for trace, active samples.</p>
 <p style="color: red;">↖ Side hole</p>	<p>G1544-80730 G1544-80700</p>	<p>Direct connect liners, single and dual taper, deactivated. Capillary column press fits into liner end, eliminating sample exposure to inlet. Ultimate protection for trace, active samples. Side hole permits use with EPC.</p>

Splitless Liner Maintenance

- Liners become contaminated with use, collecting non-volatiles, salts, excess reagents, etc., or become damaged/cracked.
- Should inspect and replace liners often.
- Handle with gloves and forceps.
- Insert into or remove liners only from cool injection ports.
- Replacing with a new liner is recommended to ensure reproducibility



Splitless Liner Maintenance (contd.)

Advantages of cleaning liners yourself:

- **Reduced cost**

Disadvantages:

- **Time-consuming**
- **Liners with special features (glass wool, cup, etc.) are difficult to clean**
- **Reproducibility of liner is compromised**
- **Removing or inserting glass wool may create significant active sites in glass**



Best advice, keep a supply of new liners on-hand!

Splitless Liner Troubleshooting

- Many chromatographic problems are blamed on the column.
- Often, a dirty liner is the culprit.

Symptoms include:

Poor peak shape

Irregular baselines

Poor resolution

Poor response

Do liner types really matter?

They do, especially for active compounds like:

- phenols
- organic acids
- pesticides
- amines
- drugs of abuse, etc.




Phenols, for example....in a separation of EPA method 8270 compounds

Cool On-Column-FID Injection of 11 Analyte Test Mix

From "Improvements in the Agilent 6890/5973 GC/MSD System for Use with USEPA Method 8270", Agilent Application Note 5988-3072EN

- | | | | | | |
|---|----------------------------|----|-----------------------|--------|--------------------|
| 1 | N-Nitrosodimethylamine | 7 | Pentachlorophenol | ISTD 1 | Dichlorobenzene-d4 |
| 2 | Aniline | 8 | Benzidine | ISTD 2 | Naphthalene-d8 |
| 3 | 2,4-Dinitrophenol | 9 | 3,3-Dichlorobenzidine | ISTD 3 | Acenaphthene-d10 |
| 4 | 4-Nitrophenol | 10 | Benzo(b)fluoranthene | ISTD 4 | Phenanthrene-d10 |
| 5 | 4,6-Dinitro-2-methylphenol | 11 | Benzo(k)fluoranthene | ISTD 5 | Chrysene-d12 |
| 6 | 4-Aminobiphenyl | | | ISTD 6 | Perylene-d12 |



Splitless Inlet Liners Tested



Hole for EPC

5062-3587 Single-taper, deactivated, with glass wool

5181-3316 Single-taper, deactivated (open top)

5181-3315 Dual-taper, deactivated (closed top)

G1544-80730 Direct Connect, single-taper, deactivated

G1544-80700 Direct Connect, Dual-taper, deactivated

Vendor X Unknown proprietary deactivation

Liner Comparison

2,4-Dinitrophenol Response Factors

Experimental:

Agilent 6890 with FID

Column = HP-5MS 30m x 0.25mm x 0.5 μ m


Compared COC to various liners

0.75 min Splitless time, 3mL/min column flow

Oven: Temp programmed per 8270 method

Inj. 250°C, Det. 300°C, Sample: 1 μ L 8270 mix

160 80 20 5 ng injected



Splitless Liner Conclusions

Agilent inlet liners can be used with a broad range of samples and analytes and chromatographic response depends heavily on liner type.


To choose a liner, first consider:

- **Type of inlet in your GC**
- **Concentration and type of sample**
 - high conc. - use Split
 - trace analytes - use Splitless or PTV
 - broad range - use Split/Splitless or PTV - general purpose
 - heat-sensitive and high boiling point compounds - use On-Column or PTV

Splitless Liner Conclusions (contd.)


Next, consider...

- **Sample size, solvent, cleanliness, and potential analyte activity** - helps to choose special liner features (cup, wool, taper, etc.) and liner volume that are necessary for your application.



Splitless Liner Conclusions (contd.)

- Finally, **optimize chromatographic conditions** for the best separation.
- Remember to **check liner condition often** and replace when necessary to minimize downtime.
- **Good chromatography starts with the inlet. Choose the correct liner for your application.**



Liner Conclusions (contd.)

Flip Top for Split/Splitless injection ports

- 30 sec liner change out
- No more hunting for that “funny looking” wrench!
- Saves fingers from getting burned. Increases instrument up time



<http://www.chem.agilent.com/Scripts/PDS.asp?IPage=12224>

Liner Conclusions (contd.)

7890 Turn Top Inlet System



Root Causes of Inlet Performance Degradation, and Consequences

Accumulation of Sample Residues

- Loss of response, tailing on active analytes, split vent trap fouling and inaccurate EPC flow control

Accumulation of Consumables wear particles

- Same as Accumulation of Sample Residues, plus “bleed peaks”


Leak in Septum Nut, Septum

- Damage to O₂ sensitive detectors, irreversible damage to column

Non-Optimized Set-up

- O-ring, Gold Seal, Ferrules, Column Nuts
- Faster inlet performance degradation between maintenance sessions

Inlet Considerations



Septa

O-Rings


Liner Type

Gold Seal

Column Ferrules


Other Considerations:
Temperature
Gas flow rates

Example Of Gross Contamination




DB-624, 30 meter megabore
Temperature program // 35°C, hold 1.50 min // 30°/min to 65°C,
hold 15 min // 20°/min to 260°, hold 50 min

Cleaning the 6890 Split/Splitless Injector




Same Column After Inlet And Column Maintenance



*Temperature program // 35°C, hold for 1.50 min //
30°/min to 65°C, hold 15 min // 20°/min to 260°C for 5 min

Split Vent Trap

What is it???



Leak in Septum




Using septa beyond lifetime/temperature recommendations.

- “Use environments” that decrease lifetime include manual injections, wrong syringe tip type, larger gauge syringes, non-Agilent Autosamplers (Agilent’s are precisely aligned).
- Septum Type and Syringe Needle type mating are essential to minimizing leak rate.

Tips to Maximize Septum Life, Minimize Septum Leaks

- Use Agilent Gold Standard, 23-26 gauge, HP Point taper syringes. The point style cores septa significantly less when used with CenterGuide Septa. Taper minimizes septum coring/wear.



HP-Point Style


- Use Agilent CenterGuide Septa. The molded hole minimizes septa coring, counter-intuitive, but true.

Solid Septum


CenterGuide Septum




High-Temperature Septa Without CenterGuide: Major Coring Before 100 Autoinjections



Agilent BTO Septa With CenterGuide: Very Little Coring Even After 700 Autoinjections




Leaks Due to Septum Nut



- With repeated use, conical needle guide gets worn, out of round, and needs replacement as septum can begin to “bulge” out, especially with excessive tightening,
- Septa fail faster because needle is not guided with as much precision.
- Under or Over tightening—tighten nut until c-clamp on top stops turning, then $\frac{1}{2}$ to $\frac{3}{4}$ turn more.
- Non-Agilent septa may be too thin, too thick, or out of round like die-cut septa and may not seal as well.
- “Use Environments” that decrease lifetime, like using non-Agilent Autosamplers (ours are precisely aligned), manual injection, larger gauge syringes
- Replace septum nut annually for peace of mind.

Other Benefits of Agilent's "Centerguide" Septa

- Packaging eliminates contamination of septa,
 - “first is as good as the last”
- Less Strain on Syringe compared to solid septa
- Bleed/Temperature Optimized, (to 400C, [trace analysis](#)), p/n 5183-4757
- Advanced Green, (to 350C, [good for general purpose](#)), p/n 5183-4759
- Long Life, (to 350C, [more injections before failure](#)), p/n 5183-4761
- Above are 50 packs, 100 packs also available.



Septa vs GC Column Costs

- Typical cost of 1 Premium Septum (list), **\$1.25**
- Typical cost of 1 GC Column, 30 m x 0.25 mm ID, **\$450.**
- No accurate leak rate detector at sub 1 mL/min flow rates.
- “Don’t step over a dollar to pick up a dime!”
- Proactively change inlet septa.

Examples of Non-Optimized Operation




Typical Cause—Re-use and mis-installation.



- Leak from O-ring, Gold Seal, ferrules, column nuts
- O-rings are elastomer compression fittings designed for one use, not perfectly elastic.
- Gold seals are designed for one use, knife edge cuts into gold layer giving leak tight seal w/o shrinkage or potential organic contaminants from polyimide out-gassing/degradation.
- Re-using could result in overlap in seal rings, resulting in a leak.



Contamination of system by residue on fingers during column installation



Procedure:

- (1) Held French fry for 5 seconds.
- (2) Fingertip was wiped with paper towel to remove as much of the offending material as possible.
- (3) Lightly touched the part of the column sticking up above the ferrule.
- (4) Installed column into injector.
- (5) Set oven temperature to 40C.
- (6) Started oven temperature program as soon as oven reached 40C.

Agilent J&W Scientific Technical Support

800-227-9770 (phone: US & Canada)*

* *Select option 3, then 3, then 1.*

866-422-5571 (fax)

www.agilent.com/chem

GC-Column-Support@Agilent.com



GC Inlet Resource Guide

Publication Number: 5988-3466ENUS

